

Regenerative Research
info@regenerativeresearch.co



Project 145125
Lab # 202920
Date Rec'd 3/16/2026
Report Issued 3/31/2026

Certificate of Analysis



Compound GLOW
Label Claim 70 mg
Batch/Lot GW26-02

<u>Analyte</u>	<u>Result</u>	<u>Units</u>	<u>LOQ</u>	<u>% of Label</u>	<u>Method</u>	<u>Date</u>	<u>CAS</u>
Peptide Analysis							
Chromatographic purity	>99.9	%	0.5		HPLC-UV/MS	3/31/2026	
BPC-157	9.02	mg	0.5	90.2	HPLC-UV-MS	3/31/2026	137525-51-0
Mass Spectral Match	ID Confirmed				HPLC-UV-MS	3/31/2026	
GHK-Cu	59.4	mg	0.5	118.8	HPLC-UV-MS	3/31/2026	49557-75-7
Mass Spectral Match	ID Confirmed				HPLC-UV-MS	3/31/2026	
Thymosin β 4	11.2	mg	0.5	112.4	HPLC-UV-MS	3/31/2026	77591-33-4
Mass Spectral Match	ID Confirmed				HPLC-UV-MS	3/31/2026	
Microbiology							
Endotoxins	<0.5	EU/mL	0.5		USP 85	3/30/2026	

The data presented are from the analysis of the sample shown and meet Krause Analytical internal quality assurance criteria unless otherwise flagged.
Methods shown reference current Krause Analytical SOPs
ND - Not detected LOQ - limit of quantification
All values reported on a per vial basis unless otherwise noted
Copyright © 2026 Krause Analytical LLC. All rights reserved.

Respectfully submitted,

Mark C. Krause
Laboratory Director


**krause analytical**

Chromatogram


Response_

021-2501.D\DAD1A (-)

Sample: 202920




Mass spectrum/Reference Spectrum




**krause analytical**


202920 332 (6.136)

Scan ES+
8.23e5

BPC-157 18 (0.333)

Scan ES+
4.69e6

202920 325 (6.007)

Scan ES+
2.55e5

Thymosin 17 (0.314)

Scan ES+
7.88e5